

PCAL End Station Measurement Trends

TXWS

Date None $\pm$ SD_None

D20221213	-0.9283 $\pm$ 0.0005
D20230124	-0.9286 $\pm$ 0.0005
D20230221	-0.9291 $\pm$ 0.0004
D20230307	-0.9289 $\pm$ 0.0004
D20230328	-0.9282 $\pm$ 0.0004
D20230523	-0.9276 $\pm$ 0.0005
D20230725	-0.9281 $\pm$ 0.0004
D20230926	-0.9285 $\pm$ 0.0004
D20231106	-0.9289 $\pm$ 0.0004
D20231107	-0.9286 $\pm$ 0.0004
D20240109	-0.9288 $\pm$ 0.0004
D20240319	-0.9269 $\pm$ 0.0005
D20240528	-0.9269 $\pm$ 0.0004
D20240718	-0.9270 $\pm$ 0.0004
D20240813	-0.9275 $\pm$ 0.0004
D20241025	-0.9282 $\pm$ 0.0004
D20241210	-0.9277 $\pm$ 0.0004
D20250204	-0.9280 $\pm$ 0.0004
D20250429	-0.9275 $\pm$ 0.0004
D20250722	-0.9277 $\pm$ 0.0005
D20250923	-0.9278 $\pm$ 0.0004
D20260203	-0.9278 $\pm$ 0.0004

Description:

Tx/WS ratio:

This ratio, α_{TW} , is the Tx to Working Standard output with the backgrounds subtracted and is made using M1 - M7

TXWS

Summary of TS / WS ratio



Mean value: -0.928032

Standard deviation: 0.000649

Standard error: 0.000142

Relative Standard error: -0.000153

TS / WS ratio



RXWS

Date None $\pm$ SD_None

D20221213	-1.3822 $\pm$ 0.0003
D20230124	-1.3820 $\pm$ 0.0004
D20230221	-1.3820 $\pm$ 0.0004
D20230307	-1.3820 $\pm$ 0.0004
D20230328	-1.3818 $\pm$ 0.0004
D20230523	-1.3797 $\pm$ 0.0004
D20230725	-1.3801 $\pm$ 0.0004
D20230926	-1.3802 $\pm$ 0.0004
D20231106	-1.3814 $\pm$ 0.0004
D20231107	-1.3819 $\pm$ 0.0003
D20240109	-1.3831 $\pm$ 0.0004
D20240319	-1.3815 $\pm$ 0.0004
D20240528	-1.3831 $\pm$ 0.0007
D20240718	-1.3828 $\pm$ 0.0006
D20240813	-1.3823 $\pm$ 0.0006
D20241025	-1.3837 $\pm$ 0.0006
D20241210	-1.3842 $\pm$ 0.0006
D20250204	-1.3846 $\pm$ 0.0005
D20250429	-1.3836 $\pm$ 0.0006
D20250722	-1.3833 $\pm$ 0.0006
D20250923	-1.3829 $\pm$ 0.0006
D20260203	-1.3848 $\pm$ 0.0005

Description:

Rx/WS ratio

This ratio, α_{RW} , is the Rx to Working Standard output with the backgrounds subtracted and is made using M7 - M9.

RXWS

Summary of RX / WS ratio


Mean value: -1.382419

Standard deviation: 0.001365

Standard error: 0.000298

Relative Standard error: -0.000215

RX / WS ratio



	rho_T
Date m5 ± SD_m5	
D20221213	7153.5622 ± 3.6907
D20230124	7155.5352 ± 3.4892
D20230221	7159.9027 ± 3.2774
D20230307	7157.9979 ± 3.2099
D20230328	7152.3889 ± 3.3863
D20230523	7145.9027 ± 3.4677
D20230725	7150.2673 ± 3.4427
D20230926	7151.8956 ± 3.3903
D20231106	7154.6400 ± 3.2434
D20231107	7152.1267 ± 3.1642
D20240109	7152.5050 ± 3.2952
D20240319	7137.9890 ± 3.5704
D20240528	7136.6224 ± 3.4316
D20240718	7138.1965 ± 3.3466
D20240813	7158.7320 ± 3.4081
D20241025	7160.9116 ± 3.2458
D20241210	7157.2848 ± 3.1588
D20250204	7152.1403 ± 3.1871
D20250429	7142.4125 ± 3.2770
D20250722	7144.7450 ± 3.8020
D20250923	7142.9465 ± 3.1430
D20260203	7145.3696 ± 3.1526

rho_T

Summary of TX Responsivity rho_T

Mean value: 7150.185201

Standard deviation: 7.351612


Standard error: 1.603810

Relative Standard error: 0.000224

Description:

TX Responsivity rho_T: $\rho_{Tx} = \alpha_{TW} * \rho_{WS}$ The responsivity of the TxPD, calculated using ρ_{WS} . Trending this parameter is useful in determining if there has been a change in the input Laser Power and for estimating how much power is incident on the ETM once corrected for optical losses inside the vacuum chamber.

TX Responsivity ρ_T




	rho_R
Date	None $\pm$ SD_None
D20221213	10651.3278 $\pm$ 2.6712
D20230124	10649.7487 $\pm$ 2.8058
D20230221	10649.6738 $\pm$ 2.8433
D20230307	10649.0726 $\pm$ 2.7307
D20230328	10648.2216 $\pm$ 2.8093
D20230523	10628.8739 $\pm$ 2.7884
D20230725	10632.5382 $\pm$ 2.7567
D20230926	10631.7067 $\pm$ 2.7108
D20231106	10639.7032 $\pm$ 3.1608
D20231107	10643.4386 $\pm$ 2.5757
D20240109	10650.7508 $\pm$ 2.8108
D20240319	10638.2119 $\pm$ 2.7655
D20240528	10649.0628 $\pm$ 5.1137
D20240718	10647.7038 $\pm$ 4.3711
D20240813	10668.9500 $\pm$ 4.6864
D20241025	10674.6112 $\pm$ 4.3725
D20241210	10679.4512 $\pm$ 4.5056
D20250204	10671.5263 $\pm$ 4.0079
D20250429	10654.3914 $\pm$ 4.5562
D20250722	10653.7194 $\pm$ 4.6842
D20250923	10646.7533 $\pm$ 4.5132
D20260203	10664.5547 $\pm$ 4.1537

rho_R
Summary of RX Responsivity rho_R
Mean value: 10651.090544
Standard deviation: 13.647751
Standard error: 2.977360
Relative Standard error: 0.000280

Description:

RX Responsivity rho_R: $p_{Rx} = \alpha_{RW} * p_{WS}$ The responsivity of the RxPD, calculated using pWS. Trending this parameter is useful in estimating how much power is incident on the ETM once corrected for optical losses inside the vacuum chamber.

RX Responsivity ρ_R





	rhoT_prime
Date	None ± SD_None
D20221213	7212.3975 ± 5.7342
D20230124	7214.4524 ± 5.5318
D20230221	7215.2736 ± 5.2105
D20230307	7213.8139 ± 5.1408
D20230328	7210.1146 ± 5.3686
D20230523	7207.4696 ± 5.5664
D20230725	7212.9695 ± 5.4961
D20230926	7213.7446 ± 5.4039
D20231106	7213.5775 ± 5.4235
D20231107	7213.3527 ± 5.0937
D20240109	7211.2352 ± 5.2532
D20240319	7193.5442 ± 5.6431
D20240528	7194.7219 ± 4.1823
D20240718	7199.6725 ± 4.0493
D20240813	7216.9582 ± 4.1529
D20241025	7217.9497 ± 3.9950
D20241210	7214.8532 ± 3.8621
D20250204	7205.8673 ± 3.8934
D20250429	7198.2687 ± 3.9789
D20250722	7203.9108 ± 4.5463
D20250923	7202.6779 ± 3.8441
D20260203	7202.5887 ± 3.8212

Description:

Tx Responsivity Corrected for Optical Efficiency:
 $\rho_{Tx}' = \rho_{Tx} * ET$
Rho_Tx^Prime is the estimated responsivity of the Tx integrating sphere if the Tx Sphere were to be located at ETM. This provides a nice sanity check for the Rx Sphere, which historically has been the main sensor for calibration.

Summary of Tx Responsivity Corrected
for Optical Efficiency
Mean value: 7208.609749
Standard deviation: 7.324413
Standard error: 1.597876
Relative Standard error: 0.000222

Tx Responsivity Corrected for Optical Efficiency




	rhoR_prime
Date	None ± SD_None
D20221213	10576.6698 ± 6.9261
D20230124	10575.0054 ± 6.8499
D20230221	10580.1815 ± 6.5500
D20230307	10578.9098 ± 6.4566
D20230328	10575.1983 ± 6.6872
D20230523	10550.2810 ± 6.9166
D20230725	10552.3120 ± 6.8061
D20230926	10552.7557 ± 6.6864
D20231106	10564.9900 ± 7.0717
D20231107	10565.3157 ± 6.3522
D20240109	10576.2380 ± 6.5885
D20240319	10568.2745 ± 6.9523
D20240528	10575.2974 ± 6.1418
D20240718	10569.0076 ± 5.4413
D20240813	10595.1249 ± 5.7782
D20241025	10602.5180 ± 5.4958
D20241210	10606.5031 ± 5.5089
D20250204	10604.2216 ± 5.1342
D20250429	10583.9556 ± 5.5794
D20250722	10578.4528 ± 5.8748
D20250923	10570.6839 ± 5.5020
D20260203	10592.0813 ± 5.1728

Description:

Rx Responsivity Corrected for Optical Efficiency:
 $\rho_{Rx}' = \rho_{Rx} / ER$
 Rho_Rx^Prime is the estimated responsivity of the Rx integrating sphere if the Rx Sphere were to be located at ETM.

 Summary of Rx Responsivity Corrected
 for Optical Efficiency
 Mean value: 10576.998995
 Standard deviation: 15.834730
 Standard error: 3.454466
 Relative Standard error: 0.000327

Rx Responsivity Corrected for Optical Efficiency



	WS
Date	WS $\pm$ SD_WS
D20221213	-4.7051 $\pm$ 0.0000
D20230124	-4.7051 $\pm$ 0.0000
D20230221	-4.7051 $\pm$ 0.0000
D20230307	-4.7051 $\pm$ 0.0000
D20230328	-4.7051 $\pm$ 0.0000
D20230523	-4.7051 $\pm$ 0.0004
D20230725	-4.7051 $\pm$ 0.0000
D20230926	-4.7051 $\pm$ 0.0000
D20231106	-4.7051 $\pm$ 0.0000
D20231107	-4.7051 $\pm$ 0.0000
D20240109	-4.7051 $\pm$ 0.0000
D20240319	-4.7051 $\pm$ 0.0000
D20240528	-4.7010 $\pm$ 0.0000
D20240718	-4.7015 $\pm$ 0.0000
D20240813	-4.7122 $\pm$ 0.0000
D20241025	-4.7100 $\pm$ 0.0000
D20241210	-4.7105 $\pm$ 0.0000
D20250204	-4.7056 $\pm$ 0.0000
D20250429	-4.7016 $\pm$ 0.0000
D20250722	-4.7022 $\pm$ 0.0000
D20250923	-4.7004 $\pm$ 0.0000
D20260203	-4.7019 $\pm$ 0.0000

Description:


WS

WS is the Working standard Responsivity as measured from the Gold standard in the lab.

Summary of Working Standard Responsivity rho_WS

Mean value: -4.704888
Standard deviation: 0.002985
Standard error: 0.000651
Relative Standard error: -0.000138

Working Standard Responsivity rho_WS




	e_i
Date	e_i ± SD_e_i
D20221213	0.9847 ± 0.0008
D20230124	0.9849 ± 0.0008
D20230221	0.9846 ± 0.0008
D20230307	0.9846 ± 0.0009
D20230328	0.9848 ± 0.0008
D20230523	0.9842 ± 0.0009
D20230725	0.9841 ± 0.0009
D20230926	0.9838 ± 0.0008
D20231106	0.9850 ± 0.0009
D20231107	0.9845 ± 0.0008
D20240109	0.9855 ± 0.0008
D20240319	0.9867 ± 0.0008
D20240528	0.9858 ± 0.0004
D20240718	0.9850 ± 0.0004
D20240813	0.9860 ± 0.0004
D20241025	0.9864 ± 0.0004
D20241210	0.9860 ± 0.0004
D20250204	0.9873 ± 0.0004
D20250429	0.9860 ± 0.0004
D20250722	0.9861 ± 0.0005
D20250923	0.9854 ± 0.0004
D20260203	0.9862 ± 0.0004

Description:

Optical Efficiency (Inner beam): $e_i = (W_{ri}/T_{ti}) / (W_{ti}/T_{ti})$
 This parameter is the ratio of the Working Standard (WS) output in both the Rx module and Tx module with only the inner beam on it, normalized by the Tx Output. This parameter involves M1 and M5 and is useful for determining overall optical Efficiency.

Summary of Optical Efficiency (Inner Beam)	
Mean value:	0.985352
Standard deviation:	0.000916
Standard error:	0.000200
Relative Standard error:	0.000203

Optical Efficiency (Inner Beam)



	e_o
Date	e_o ± SD_e_o
D20221213	0.9851 ± 0.0009
D20230124	0.9849 ± 0.0008
D20230221	0.9871 ± 0.0008
D20230307	0.9868 ± 0.0007
D20230328	0.9856 ± 0.0009
D20230523	0.9840 ± 0.0008
D20230725	0.9835 ± 0.0008
D20230926	0.9843 ± 0.0008
D20231106	0.9847 ± 0.0008
D20231107	0.9839 ± 0.0008
D20240109	0.9844 ± 0.0008
D20240319	0.9848 ± 0.0009
D20240528	0.9843 ± 0.0005
D20240718	0.9832 ± 0.0004
D20240813	0.9841 ± 0.0005
D20241025	0.9844 ± 0.0005
D20241210	0.9845 ± 0.0004
D20250204	0.9852 ± 0.0004
D20250429	0.9854 ± 0.0004
D20250722	0.9835 ± 0.0005
D20250923	0.9839 ± 0.0004
D20260203	0.9844 ± 0.0004


Description:

Optical Efficiency (Outer beam): $e_o = (W_{ro}/T_{to}) / (W_{to}/T_{to})$
 This parameter is the ratio of the Working Standard (WS) output in both the Rx module and Tx module with only the outer beam on it, normalized by the Tx Output. This parameter involves M2 and M4 and is useful for determining overall optical Efficiency.

Summary of Optical Efficiency (Outer Beam)

Mean value:	0.984639
Standard deviation:	0.000962
Standard error:	0.000210
Relative Standard error:	0.000213

Optical Efficiency (Outer Beam)



	e
Date	None $\pm$ SD_None
D20221213	0.9849 $\pm$ 0.0006
D20230124	0.9849 $\pm$ 0.0006
D20230221	0.9859 $\pm$ 0.0006
D20230307	0.9857 $\pm$ 0.0005
D20230328	0.9852 $\pm$ 0.0006
D20230523	0.9841 $\pm$ 0.0006
D20230725	0.9838 $\pm$ 0.0006
D20230926	0.9841 $\pm$ 0.0006
D20231106	0.9849 $\pm$ 0.0006
D20231107	0.9842 $\pm$ 0.0005
D20240109	0.9849 $\pm$ 0.0006
D20240319	0.9858 $\pm$ 0.0006
D20240528	0.9851 $\pm$ 0.0006
D20240718	0.9841 $\pm$ 0.0006
D20240813	0.9851 $\pm$ 0.0006
D20241025	0.9854 $\pm$ 0.0006
D20241210	0.9852 $\pm$ 0.0006
D20250204	0.9863 $\pm$ 0.0006
D20250429	0.9857 $\pm$ 0.0006
D20250722	0.9848 $\pm$ 0.0007
D20250923	0.9846 $\pm$ 0.0006
D20260203	0.9853 $\pm$ 0.0006



Description:

Overall Optical Efficiency: $e = (e_o + e_i)/2$
 This parameter is just the average efficiency of both the inner and outer beams. This uses M1 ,M2, M4, & M5 and is used to determine how much power is lost inside the vacuum chamber.

 e

Summary of Overall Optical Efficiency	
Mean value:	0.984995
Standard deviation:	0.000650
Standard error:	0.000142
Relative Standard error:	0.000144

Overall Optical Efficiency



	E_T
Date	E_T ± SD_E_T
D20221213	0.9918 ± 0.0006
D20230124	0.9918 ± 0.0006
D20230221	0.9923 ± 0.0006
D20230307	0.9923 ± 0.0005
D20230328	0.9920 ± 0.0006
D20230523	0.9915 ± 0.0006
D20230725	0.9913 ± 0.0006
D20230926	0.9914 ± 0.0006
D20231106	0.9918 ± 0.0006
D20231107	0.9915 ± 0.0005
D20240109	0.9919 ± 0.0006
D20240319	0.9923 ± 0.0006
D20240528	0.9919 ± 0.0003
D20240718	0.9915 ± 0.0003
D20240813	0.9919 ± 0.0003
D20241025	0.9921 ± 0.0003
D20241210	0.9920 ± 0.0003
D20250204	0.9925 ± 0.0003
D20250429	0.9922 ± 0.0003
D20250722	0.9918 ± 0.0003
D20250923	0.9917 ± 0.0003
D20260203	0.9921 ± 0.0003


Description:

Tx side Optical Efficiency Correction Factor: $E_T = \sqrt{e \cdot \beta}$

This parameter is used to apportion the optical loss inside the vacuum chamber to before the ETM and after the ETM, and uses the Optical Efficiency (e) along with the last in-chamber measurements (DCC- T1800107). The in-chamber measurements yield a ratio of optical efficiencies between the Tx side and Rx side, referred to as β , which we assume have not changed. Applying this correction factor corrects for the optical loss on the Tx side, providing the Rx and Tx responsivities at the ETM.

Summary of Input Side Optical Efficiency	E_T
	correction factor
Mean value:	0.991895
Standard deviation:	0.000328
Standard error:	0.000071
Relative Standard error:	0.000072

Input Side Optical Efficiency correction factor



	E_R
Date	None $\pm$ SD_None
D20221213	0.9930 $\pm$ 0.0006
D20230124	0.9930 $\pm$ 0.0006
D20230221	0.9935 $\pm$ 0.0006
D20230307	0.9934 $\pm$ 0.0006
D20230328	0.9931 $\pm$ 0.0006
D20230523	0.9926 $\pm$ 0.0006
D20230725	0.9925 $\pm$ 0.0006
D20230926	0.9926 $\pm$ 0.0006
D20231106	0.9930 $\pm$ 0.0006
D20231107	0.9927 $\pm$ 0.0005
D20240109	0.9930 $\pm$ 0.0006
D20240319	0.9934 $\pm$ 0.0006
D20240528	0.9931 $\pm$ 0.0003
D20240718	0.9926 $\pm$ 0.0003
D20240813	0.9931 $\pm$ 0.0003
D20241025	0.9932 $\pm$ 0.0003
D20241210	0.9932 $\pm$ 0.0003
D20250204	0.9937 $\pm$ 0.0003
D20250429	0.9934 $\pm$ 0.0003
D20250722	0.9929 $\pm$ 0.0003
D20250923	0.9929 $\pm$ 0.0003
D20260203	0.9932 $\pm$ 0.0003


Summary of Output Side Optical Efficiency correction factor	
Mean value:	0.993044
Standard deviation:	0.000328
Standard error:	0.000072
Relative Standard error:	0.000072

Description:

Rx side Optical Efficiency Correction Factor: $E_R = \sqrt{e/ \beta}$

This parameter is used to apportion the optical loss inside the vacuum chamber to before the ETM and after the ETM, and uses the Optical Efficiency (e) along with the last in-chamber measurements (DCC- T1800107). The in-chamber measurements yield a ratio of optical efficiencies between the Tx side and Rx side, referred to as β , which we assume have not changed. Applying this correction factor corrects for the optical loss on the Rx side, providing the Rx and Tx responsivities at the ETM.

Output Side Optical Efficiency correction factor



	PowerImbalance
Date	None $\pm$ SD_None
D20221213	0.9855 $\pm$ 0.0005
D20230124	0.9870 $\pm$ 0.0005
D20230221	0.9897 $\pm$ 0.0004
D20230307	0.9886 $\pm$ 0.0004
D20230328	0.9862 $\pm$ 0.0004
D20230523	0.9858 $\pm$ 0.0005
D20230725	0.9883 $\pm$ 0.0005
D20230926	0.9880 $\pm$ 0.0004
D20231106	0.9867 $\pm$ 0.0004
D20231107	0.9867 $\pm$ 0.0004
D20240109	0.9862 $\pm$ 0.0004
D20240319	0.9870 $\pm$ 0.0005
D20240528	0.9855 $\pm$ 0.0005
D20240718	0.9858 $\pm$ 0.0004
D20240813	0.9860 $\pm$ 0.0004
D20241025	0.9840 $\pm$ 0.0008
D20241210	0.9841 $\pm$ 0.0007
D20250204	0.9811 $\pm$ 0.0008
D20250429	0.9848 $\pm$ 0.0008
D20250722	0.9842 $\pm$ 0.0009
D20250923	0.9841 $\pm$ 0.0007
D20260203	0.9825 $\pm$ 0.0007

Description:

Power Imbalance:
The Power imbalance is a measurement of Inner Beam Power divided by Outer Beam Power, as measured by the Working Standard inside the TX module. M1/M2.

	PowerImbalance
Summary of Power Imbalance	
Mean value:	0.985804
Standard deviation:	0.002006
Standard error:	0.000438
Relative Standard error:	0.000444

Power Imbalance

